Mnemonic's

AC0VXG	mains supply 0V X-ray generator
AC230VL1	mains supply 230V AC phase 1
AC230VL2	mains supply 230V AC phase 2
GND12V	ground +12V supply
MNEMOF	mains power emergency off
MNON/	mains on
PO12V	+12V supply
PWONNG	DC supply relay power on -
PWONPO	DC supply relay power on +
RDMNON/	ready mains power on
RESETSW/	system reset
SIPHIO	single phase intijicotion

Wall junction box
MEX
Mains supply
3~400V
50/60Hz
F1 25A
X3201
N
PE
K1
K2
R1 47R
R2 47R
R3 47R
EQX1101 (E2QX1101)
EQX1102 (E2QX1102)
EQX1103 (E2QX1103)
EQ200 X12 (E2Q200 X12)
Discharging circuit
EQ200 X11 (E2Q200 X11)
F2 10A
F12 10A
X23
X22
NX21
Output 3
Imaging
Output 2
II/TV
GECO
SYSCO
Output 1
Geometry
K11
EN
X3:1
X3:3
T1
230V~
X3:2
H1 gr
low voltage control
100
Power on circuit
S1
V1
X1:6
X1:7
X1:2
X1:1
X2:1
X2:2
X1:9
X1:5
X1:4
T1
T2
X2402
X2403
A1
A2
RC1
F11 5A
F3 5A
X3102
X3101
X1102
WG 51
WG 57
EM off
EZ X13:3
EZ X13:2
EZ X13:1
EZ X47:6
EZ X47:7
EZ X47:2
EZ X47:1
EZ X47:9
EZ X47:5
EZ X47:4
EZ X4:1
EZ X4:2
GND 12V
RDMNON/
MNON/
PO26V
GND26V
PWONNG
SIPHIO
PWONPO
EZ X14:1
EZ X14:2
EZ X15:1
EZ X15:2
EZ X46:15
EZ X46:14
EZ X46:6
EZ X20:1
EZ X20:2
X10:11
X10:4
X10:3
X10:1
X10:15
Logic
off
on
YM
M1 RoCo
GS1:1
M cooler
H.V. generator
G200 X2
Oil circulation pump
H.V. generator
RoCo LS
RoCo HS
WG
Tube extension (option)
EZ119
mA control
EZ102
low voltage supply
EWN
second low voltage supply (option)
Adaptation
24V-/230V~
EZ130
kV control
see Z1 :H.V. generator R/F
CB100 Cockpit
Power supply

A2/A3 02-04-30 Schr./Kö

© Philips Medical Systems